

ZytoLight® SPEC MYB Dual Color Break Apart Probe



Background


The ZytoLight® SPEC MYB Dual Color Break Apart Probe (PL100) is intended to be used for the qualitative detection of translocations involving the human MYB gene at 6q23.3 in cytologic or formalin-fixed, paraffin-embedded specimens, such as adenoid cystic carcinoma (ACC), by fluorescence *in situ* hybridization (FISH). The probe is intended to be used in combination with ZytoLight® FISH Implementation Kits (Prod. No. Z-2028-5/-20, or Z-2099-20).

The product is intended for professional use only. All tests using the product should be performed in a certified, licensed anatomic pathology laboratory under the supervision of a pathologist/human geneticist by qualified personnel. The probe is intended to be used as an aid to the differential diagnosis of ACC and therapeutic measures should not be initiated based on the test result alone.


Probe Description

The ZytoLight® SPEC MYB Dual Color Break Apart Probe is composed of:

- ZyGreen (excitation 503 nm/emission 528 nm) labeled polynucleotides (~10.0 ng/μl), which target sequences mapping in 6q23.2-q23.3** (chr6:134,840,690-135,483,752) proximal to the MYB breakpoint region.
- ZyOrange (excitation 547 nm/emission 572 nm) labeled polynucleotides (~4.5 ng/μl), which target sequences mapping in 6q23.3** (chr6:135,728,667-136,390,142) distal to the MYB breakpoint region.
- Formamide based hybridization buffer




Ideogram of chromosome 6 indicating the hybridization locations.




SPEC MYB Probe map (not to scale).

Results

In an interphase nucleus lacking a translocation involving the 6q23.2-q23.3 band, two orange/green fusion signals are expected representing two normal (non-rearranged) 6q23.2-q23.3 loci. A signal pattern consisting of one orange/green fusion signal, one orange signal, and a separate green signal indicates one normal 6q23.2-q23.3 locus and one 6q23.2-q23.3 locus affected by a translocation.



SPEC MYB Dual Color Break Apart Probe hybridized to normal interphase cells as indicated by two orange/green fusion signals per nucleus.



Adenoid cystic carcinoma tissue section with translocation affecting the 6q23.3 locus as indicated by one orange/green fusion (non-rearranged) signal, one orange signal, and one separate green signal.

Prod. No.	Product	Label	Tests* (Volume)
Z-2143-50	ZytoLight SPEC MYB Dual Color Break Apart Probe	●/●	5 (50 μl)
Z-2143-200	ZytoLight SPEC MYB Dual Color Break Apart Probe	●/●	20 (200 μl)
Related Products			
Z-2028-5	ZytoLight FISH-Tissue Implementation Kit Incl. Heat Pretreatment Solution Citric, 150 ml; Pepsin Solution, 1 ml; Wash Buffer SSC, 210 ml; 25x Wash Buffer A, 50 ml; DAPI/DuraTect-Solution, 0.2 ml		5
Z-2028-20	ZytoLight FISH-Tissue Implementation Kit Incl. Heat Pretreatment Solution Citric, 500 ml; Pepsin Solution, 4 ml; Wash Buffer SSC, 560 ml; 25x Wash Buffer A, 100 ml; DAPI/DuraTect-Solution, 0.8 ml		20
Z-2099-20	ZytoLight FISH-Cytology Implementation Kit Incl. Cytology Pepsin Solution, 4 ml; 20x Wash Buffer TBS, 50 ml; 10x MgCl ₂ , 50 ml; 10x PBS, 50 ml; Cytology Stringency Wash Buffer SSC, 500 ml; Cytology Wash Buffer SSC, 500 ml; DAPI/DuraTect-Solution, 0.8 ml		20

* Using 10 μl probe solution per test. labeled products are only available in certain countries. All other countries research use only! Please contact your local dealer for more information.

**According to Human Genome Assembly GRCh37/hg19